

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
 Data File : VX028599.D
 Acq On : 10 May 2022 17:09
 Operator : JC/MD
 Sample : N2745-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-31S

Quant Time: May 11 07:02:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	402836	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	696841	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	686722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	327277	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	272726	51.923	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.840%	
35) Dibromofluoromethane	5.391	113	254648	52.108	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.220%	
50) Toluene-d8	8.653	98	941466	50.159	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.320%	
62) 4-Bromofluorobenzene	11.079	95	349425	48.689	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.380%	
Target Compounds						Qvalue
16) Acetone	2.379	43	43260	25.713	ug/l	99
20) Methylene Chloride	2.794	84	12971	3.328	ug/l	94
43) Isopropyl Acetate	6.348	43	28903	2.746	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
Data File : VX028599.D
Acq On : 10 May 2022 17:09
Operator : JC/MD
Sample : N2745-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-31S

Quant Time: May 11 07:02:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

