

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
 Data File : VX028600.D
 Acq On : 10 May 2022 17:32
 Operator : JC/MD
 Sample : N2745-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-31D

Quant Time: May 11 07:02:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	383437	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	662006	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	656867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	315934	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	265982	53.201	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.400%	
35) Dibromofluoromethane	5.391	113	245923	52.971	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.940%	
50) Toluene-d8	8.653	98	913213	51.214	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.420%	
62) 4-Bromofluorobenzene	11.079	95	340737	49.977	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.960%	
Target Compounds						Qvalue
16) Acetone	2.380	43	45511	28.567	ug/l	93
20) Methylene Chloride	2.788	84	10510	2.833	ug/l	97
43) Isopropyl Acetate	6.348	43	28491	2.849	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
Data File : VX028600.D
Acq On : 10 May 2022 17:32
Operator : JC/MD
Sample : N2745-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-31D

Quant Time: May 11 07:02:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

