

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX051022\  
 Data File : VX028605.D  
 Acq On : 10 May 2022 19:28  
 Operator : JC/MD  
 Sample : N2749-14  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TP-5

Quant Time: May 11 07:04:29 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 19 13:38:04 2022  
 Response via : Initial Calibration

| Compound                    | R.T. QIon |       | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|-----------|-------|----------|----------|------------|----------|
| -----                       |           |       |          |          |            |          |
| Internal Standards          |           |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.556     | 168   | 389832   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763     | 114   | 667959   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055    | 117   | 663858   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024    | 152   | 328717   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |           |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.958     | 65    | 269834   | 53.086   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 61 - 141 | Recovery | = 106.180% |          |
| 35) Dibromofluoromethane    | 5.391     | 113   | 247635   | 52.864   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 69 - 133 | Recovery | = 105.720% |          |
| 50) Toluene-d8              | 8.653     | 98    | 912296   | 50.707   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 65 - 126 | Recovery | = 101.420% |          |
| 62) 4-Bromofluorobenzene    | 11.079    | 95    | 346544   | 50.375   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 58 - 135 | Recovery | = 100.760% |          |
| Target Compounds            |           |       |          |          |            |          |
|                             |           |       |          |          |            | Qvalue   |
| 16) Acetone                 | 2.380     | 43    | 41403    | 25.415   | ug/l       | 95       |
| 20) Methylene Chloride      | 2.788     | 84    | 29212    | 7.746    | ug/l       | 96       |
| 43) Isopropyl Acetate       | 6.348     | 43    | 27412    | 2.717    | ug/l       | 97       |
| -----                       |           |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX051022\  
Data File : VX028605.D  
Acq On : 10 May 2022 19:28  
Operator : JC/MD  
Sample : N2749-14  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TP-5

Quant Time: May 11 07:04:29 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 19 13:38:04 2022  
Response via : Initial Calibration

