

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
 Data File : VX028607.D
 Acq On : 10 May 2022 20:15
 Operator : JC/MD
 Sample : N2749-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-7

Quant Time: May 11 07:05:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	395637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	675903	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	681450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	326408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	268790	52.105	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.200%	
35) Dibromofluoromethane	5.391	113	251998	53.164	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.320%	
50) Toluene-d8	8.653	98	929817	51.073	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.140%	
62) 4-Bromofluorobenzene	11.079	95	345754	49.670	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.340%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	40985	24.754	ug/l	98
20) Methylene Chloride	2.794	84	33057	8.636	ug/l	98
43) Isopropyl Acetate	6.348	43	27610	2.704	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
Data File : VX028607.D
Acq On : 10 May 2022 20:15
Operator : JC/MD
Sample : N2749-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-7

Quant Time: May 11 07:05:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

