

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
 Data File : VX028609.D
 Acq On : 10 May 2022 21:01
 Operator : JC/MD
 Sample : PB144702ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144702ZHE#10

Quant Time: May 11 07:05:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	379325	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	664043	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	666234	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	316944	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	267632	54.111	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.220%	
35) Dibromofluoromethane	5.391	113	248334	53.326	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.660%	
50) Toluene-d8	8.653	98	919799	51.425	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.860%	
62) 4-Bromofluorobenzene	11.079	95	342565	50.091	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 100.180%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	42194	26.684	ug/l	92
20) Methylene Chloride	2.788	84	36662	9.990	ug/l	97
43) Isopropyl Acetate	6.342	43	30047	2.995	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
Data File : VX028609.D
Acq On : 10 May 2022 21:01
Operator : JC/MD
Sample : PB144702ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144702ZHE#10

Quant Time: May 11 07:05:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

