

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051023\
 Data File : VX035579.D
 Acq On : 10 May 2023 12:02
 Operator : JC/MD
 Sample : VX0510WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0510WBL01

Quant Time: May 11 04:15:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	303626	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	492777	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	456660	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	232591	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	156308	42.671	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.340%
35) Dibromofluoromethane	5.385	113	154485	51.078	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.160%
50) Toluene-d8	8.647	98	569056	49.876	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.760%
62) 4-Bromofluorobenzene	11.079	95	227264	52.950	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051023\
Data File : VX035579.D
Acq On : 10 May 2023 12:02
Operator : JC/MD
Sample : VX0510WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0510WBL01

Quant Time: May 11 04:15:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

