

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051023\
 Data File : VX035592.D
 Acq On : 10 May 2023 17:33
 Operator : JC/MD
 Sample : 02719-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-01_20230509

Quant Time: May 11 04:29:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	337456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	539083	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	498506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	248911	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	169373	41.603	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.200%
35) Dibromofluoromethane	5.385	113	165529	50.028	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.060%
50) Toluene-d8	8.646	98	620052	49.677	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.079	95	244433	52.058	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.120%
Target Compounds						
16) Acetone	2.398	43	3795	2.146	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051023\
Data File : VX035592.D
Acq On : 10 May 2023 17:33
Operator : JC/MD
Sample : 02719-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-01_20230509

Quant Time: May 11 04:29:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

