

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051023\
 Data File : VX035594.D
 Acq On : 10 May 2023 18:19
 Operator : JC/MD
 Sample : 02743-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DRUM-A-WATER

Quant Time: May 11 04:29:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	326915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	526513	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	486366	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	237181	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	167414	42.447	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.900%
35) Dibromofluoromethane	5.385	113	160502	49.667	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.340%
50) Toluene-d8	8.647	98	603044	49.468	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.940%
62) 4-Bromofluorobenzene	11.079	95	238732	52.058	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.120%
Target Compounds						
16) Acetone	2.398	43	3990	2.329	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051023\
Data File : VX035594.D
Acq On : 10 May 2023 18:19
Operator : JC/MD
Sample : 02743-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DRUM-A-WATER

Quant Time: May 11 04:29:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

