

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX051119\
 Data File : VX009515.D
 Acq On : 11 May 2019 11:12
 Operator : JC/SP
 Sample : VX0511WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0511WBL01

Quant Time: May 12 02:11:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	196968	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	310743	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	266609	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	110534	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	128343	47.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.46%	
35) Dibromofluoromethane	5.50	113	94976	48.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.24%	
50) Toluene-d8	8.71	98	388056	49.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.40%	
62) 4-Bromofluorobenzene	11.13	95	126055	43.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.98%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX051119\
Data File : VX009515.D
Acq On : 11 May 2019 11:12
Operator : JC/SP
Sample : VX0511WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX0511WBL01

Quant Time: May 12 02:11:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

