

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX051119\
 Data File : VX009524.D
 Acq On : 11 May 2019 14:44
 Operator : JC/SP
 Sample : K2799-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Manual Integrations
 APPROVED

MMDadoda
 5/13/2019 4:27:20 PM

Quant Time: May 12 02:28:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	185856	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	291050	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	250678	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	103951	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	120416	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
35) Dibromofluoromethane	5.49	113	89093	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
50) Toluene-d8	8.71	98	361421	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
62) 4-Bromofluorobenzene	11.13	95	116944	43.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.14%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	162437	71.241	ug/l	96
12) 1,1-Dichloroethene	2.37	96	3088	1.771	ug/l	96
16) Acetone	2.44	43	4194	3.078	ug/l	95
21) trans-1,2-Dichloroethene	3.17	96	2109	1.115	ug/l #	75
27) cis-1,2-Dichloroethene	4.60	96	586870	259.977	ug/l	89
44) Trichloroethene	7.21	130	35999	17.789	ug/l	96
64) Tetrachloroethene	9.34	164	5423	3.154	ug/l	96
65) Chlorobenzene	10.14	112	1853	0.370	ug/l	96
87) 1,3-Dichlorobenzene	12.02	146	1370	0.410	ug/l	93
88) 1,4-Dichlorobenzene	12.10	146	765m	0.227	ug/l	
91) 1,2-Dichlorobenzene	12.39	146	7397	2.171	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051119\
Data File : VX009524.D
Acq On : 11 May 2019 14:44
Operator : JC/SP
Sample : K2799-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Manual Integrations
APPROVED

MMDadoda
5/13/2019 4:27:20 PM

Quant Time: May 12 02:28:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

