

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051119\
 Data File : VX009526.D
 Acq On : 11 May 2019 15:30
 Operator : JC/SP
 Sample : K2799-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW-6

Quant Time: May 12 02:31:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	183123	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	290982	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	249394	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	100913	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119770	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
35) Dibromofluoromethane	5.50	113	88497	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
50) Toluene-d8	8.71	98	360531	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
62) 4-Bromofluorobenzene	11.13	95	115301	42.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.94%	
Target Compounds						
16) Acetone	2.44	43	5503	4.100	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051119\
Data File : VX009526.D
Acq On : 11 May 2019 15:30
Operator : JC/SP
Sample : K2799-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW-6

Quant Time: May 12 02:31:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

