

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051119\
 Data File : VX009528.D
 Acq On : 11 May 2019 16:17
 Operator : JC/SP
 Sample : K2799-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: May 12 02:34:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	185325	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	295382	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	255490	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	109263	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	120771	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
35) Dibromofluoromethane	5.50	113	89457	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
50) Toluene-d8	8.71	98	368832	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
62) 4-Bromofluorobenzene	11.13	95	119866	44.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.00%	
Target Compounds						
16) Acetone	2.44	43	6008	4.423	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051119\
Data File : VX009528.D
Acq On : 11 May 2019 16:17
Operator : JC/SP
Sample : K2799-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: May 12 02:34:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

