

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051120\
 Data File : VX016144.D
 Acq On : 11 May 2020 13:26
 Operator : JC/SP
 Sample : L2552-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z4-DUP-0553-200507

Quant Time: May 13 09:09:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	291757	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	476811	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	466512	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	242310	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	177668	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
35) Dibromofluoromethane	5.47	113	141238	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
50) Toluene-d8	8.70	98	589456	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
62) 4-Bromofluorobenzene	11.12	95	236105	53.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.66%	
Target Compounds						
12) 1,1-Dichloroethene	2.36	96	1639	0.540	ug/l	91
16) Acetone	2.42	43	2553	1.490	ug/l #	82
19) Methyl tert-butyl Ether	3.17	73	12430	1.189	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	14293	3.868	ug/l	82
44) Trichloroethene	7.19	130	55180	11.310	ug/l	96
64) Tetrachloroethene	9.32	164	9851	1.720	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051120\
Data File : VX016144.D
Acq On : 11 May 2020 13:26
Operator : JC/SP
Sample : L2552-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z4-DUP-0553-200507

Quant Time: May 13 09:09:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

