

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051120\
 Data File : VX016154.D
 Acq On : 11 May 2020 17:12
 Operator : JC/SP
 Sample : PB128761ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB128761ZHE#04

Quant Time: May 12 08:30:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	288316	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	479495	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	473004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	242297	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	174903	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
35) Dibromofluoromethane	5.47	113	143204	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
50) Toluene-d8	8.70	98	592527	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
62) 4-Bromofluorobenzene	11.12	95	237702	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
Target Compounds						
16) Acetone	2.42	43	25812	15.248	ug/l	95
18) Methyl Acetate	2.75	43	4193	1.007	ug/l	96
20) Methylene Chloride	2.83	84	3670	1.052	ug/l	97
43) Isopropyl Acetate	6.42	43	95927	10.990	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051120\
Data File : VX016154.D
Acq On : 11 May 2020 17:12
Operator : JC/SP
Sample : PB128761ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB128761ZHE#04

Quant Time: May 12 08:30:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

