

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051120\
 Data File : VX016155.D
 Acq On : 11 May 2020 17:35
 Operator : JC/SP
 Sample : PB128761ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB128761ZHE#05

Quant Time: May 12 08:31:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	290984	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	483152	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	480154	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	241684	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	178586	46.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.28%	
35) Dibromofluoromethane	5.47	113	144691	47.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.54%	
50) Toluene-d8	8.70	98	599222	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	11.12	95	238985	53.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.54%	
Target Compounds						
16) Acetone	2.42	43	25989	15.212	ug/l	98
18) Methyl Acetate	2.75	43	4333	1.031	ug/l	97
43) Isopropyl Acetate	6.42	43	94618	10.758	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051120\
Data File : VX016155.D
Acq On : 11 May 2020 17:35
Operator : JC/SP
Sample : PB128761ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB128761ZHE#05

Quant Time: May 12 08:31:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

