

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051120\
 Data File : VX016159.D
 Acq On : 11 May 2020 19:06
 Operator : JC/SP
 Sample : L2502-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PL-01-050820

Quant Time: May 12 08:38:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	293336	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	487397	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	478552	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	246249	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	183186	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
35) Dibromofluoromethane	5.47	113	148130	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
50) Toluene-d8	8.70	98	607581	51.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.14%	
62) 4-Bromofluorobenzene	11.12	95	242203	53.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.04%	
Target Compounds						
16) Acetone	2.42	43	26213	15.220	ug/l	94
20) Methylene Chloride	2.84	84	190626	53.690	ug/l	99
43) Isopropyl Acetate	6.42	43	74511	8.398	ug/l	100
95) Naphthalene	13.82	128	130447	6.808	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051120\
Data File : VX016159.D
Acq On : 11 May 2020 19:06
Operator : JC/SP
Sample : L2502-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PL-01-050820

Quant Time: May 12 08:38:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

