

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051120\
 Data File : VX016162.D
 Acq On : 11 May 2020 20:14
 Operator : JC/SP
 Sample : L2571-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BUS-LANE-1

Quant Time: May 12 08:44:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	304765	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	502828	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	490020	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	245697	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	184291	45.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.90%	
35) Dibromofluoromethane	5.47	113	148776	46.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.40%	
50) Toluene-d8	8.70	98	617862	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	11.12	95	246096	52.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.42%	
Target Compounds						
16) Acetone	2.42	43	45944	25.676	ug/l	99
18) Methyl Acetate	2.75	43	5367	1.220	ug/l	97
20) Methylene Chloride	2.84	84	12744	3.455	ug/l	96
25) 2-Butanone	4.64	43	18999	6.589	ug/l	98
43) Isopropyl Acetate	6.42	43	82747	9.040	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051120\
Data File : VX016162.D
Acq On : 11 May 2020 20:14
Operator : JC/SP
Sample : L2571-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
BUS-LANE-1

Quant Time: May 12 08:44:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

