

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051121\
 Data File : VX021890.D
 Acq On : 11 May 2021 10:47
 Operator : JC/MD
 Sample : VX0511WBL01
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK301

Quant Time: May 12 03:38:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR051021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 12 03:37:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	81554	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	71867	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	33258	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	34323	4.90	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.00%
7) Chloroethane-d5	1.672	69	27959	4.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.60%
11) 1,1-Dichloroethene-d2	2.313	63	43813	3.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.40%
20) 2-Butanone-d5	4.477	46	80159	50.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.46%
24) Chloroform-d	5.062	84	58202	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.80%
26) 1,2-Dichloroethane-d4	5.970	65	31150	5.04	ug/L	0.01
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
32) Benzene-d6	5.983	84	121898	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
36) 1,2-Dichloropropane-d6	7.312	67	38295	5.24	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.80%
41) Toluene-d8	8.653	98	111665	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
43) trans-1,3-Dichloroprop...	8.952	79	16805	5.10	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.00%
46) 2-Hexanone-d5	9.391	63	72011	51.36	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.72%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	28030	4.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.40%
66) 1,2-Dichlorobenzene-d4	12.323	152	36242	5.20	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.00%
Target Compounds						Qvalue
16) Methylene chloride	2.800	84	2024	0.25	ug/L	81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051121\
Data File : VX021890.D
Acq On : 11 May 2021 10:47
Operator : JC/MD
Sample : VX0511WBL01
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK301

Quant Time: May 12 03:38:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR051021WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 12 03:37:54 2021
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.25 ug/L
RT: 2.800 min Scan# 281
Delta R.T. 0.006 min
Lab File: VX021890.D
Acq: 11 May 2021 10:47

Instrument :
MSVOA_X
ClientSampleId :
VBLK301

Tgt Ion: 84 Resp: 2024

Ion	Ratio	Lower	Upper
84	100		
86	51.9	46.3	86.1
49	98.3	84.4	156.7

