

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051123\
 Data File : VX035610.D
 Acq On : 11 May 2023 16:17
 Operator : JC/MD
 Sample : 02691-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-504-VOC-07

Quant Time: May 11 16:47:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	185492	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	341948	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	314879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146392	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	155085	47.925	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.860%
35) Dibromofluoromethane	5.385	113	111842	48.499	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.000%
50) Toluene-d8	8.647	98	418042	49.070	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.140%
62) 4-Bromofluorobenzene	11.079	95	176229	51.375	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.760%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	44218	31.793	ug/l	95
20) Methylene Chloride	2.794	84	11705	4.644	ug/l	94
37) Ethyl Acetate	4.708	43	53816m	14.174	ug/l	
43) Isopropyl Acetate	6.336	43	217948	32.920	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051123\
Data File : VX035610.D
Acq On : 11 May 2023 16:17
Operator : JC/MD
Sample : 02691-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-504-VOC-07

Quant Time: May 11 16:47:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

