

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051123\
 Data File : VX035612.D
 Acq On : 11 May 2023 17:03
 Operator : JC/MD
 Sample : PB152715TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB152715TB

Quant Time: May 12 03:51:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	192596	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	344050	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	317992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	156159	46.477	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.960%
35) Dibromofluoromethane	5.385	113	113805	49.048	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.100%
50) Toluene-d8	8.647	98	422621	49.305	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.600%
62) 4-Bromofluorobenzene	11.079	95	175621	50.885	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.780%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	39780	27.547	ug/l	92
20) Methylene Chloride	2.788	84	9905	3.785	ug/l	91
37) Ethyl Acetate	4.715	43	59027	15.452	ug/l	98
43) Isopropyl Acetate	6.336	43	237043	35.585	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051123\
Data File : VX035612.D
Acq On : 11 May 2023 17:03
Operator : JC/MD
Sample : PB152715TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB152715TB

Quant Time: May 12 03:51:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

