

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051123\
 Data File : VX035617.D
 Acq On : 11 May 2023 19:00
 Operator : JC/MD
 Sample : 02747-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-20230510

Quant Time: May 12 03:54:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	180818	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	335239	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	309824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142122	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	151061	47.888	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.780%
35) Dibromofluoromethane	5.385	113	107784	47.674	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.340%
50) Toluene-d8	8.647	98	401764	48.103	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.200%
62) 4-Bromofluorobenzene	11.079	95	167561	49.826	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051123\
Data File : VX035617.D
Acq On : 11 May 2023 19:00
Operator : JC/MD
Sample : 02747-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-20230510

Quant Time: May 12 03:54:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

