

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX051123\  
 Data File : VX035618.D  
 Acq On : 11 May 2023 19:23  
 Operator : JC/MD  
 Sample : 02755-01 50X  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 A5560

Quant Time: May 12 03:54:23 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051123W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 11 13:57:54 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)  |
|-----------------------------|--------|-------|----------|----------|-------|-----------|
| -----                       |        |       |          |          |       |           |
| Internal Standards          |        |       |          |          |       |           |
| 1) Pentafluorobenzene       | 5.556  | 168   | 185568   | 50.000   | ug/l  | # 0.00    |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 339722   | 50.000   | ug/l  | 0.00      |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 309886   | 50.000   | ug/l  | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 138037   | 50.000   | ug/l  | 0.00      |
| System Monitoring Compounds |        |       |          |          |       |           |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 153723   | 47.485   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 94.960%   |
| 35) Dibromofluoromethane    | 5.385  | 113   | 109871   | 47.956   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 95.920%   |
| 50) Toluene-d8              | 8.647  | 98    | 410063   | 48.449   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 96.900%   |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 166636   | 48.897   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 97.800%   |
| Target Compounds            |        |       |          |          |       |           |
| 52) Toluene                 | 8.720  | 92    | 1562     | 0.245    | ug/l  | Qvalue 92 |
| -----                       |        |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX051123\  
Data File : VX035618.D  
Acq On : 11 May 2023 19:23  
Operator : JC/MD  
Sample : 02755-01 50X  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
A5560

Quant Time: May 12 03:54:23 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051123W.M  
Quant Title : SW846 8260  
QLast Update : Thu May 11 13:57:54 2023  
Response via : Initial Calibration

