

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051123\
 Data File : VX035624.D
 Acq On : 11 May 2023 21:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 12 03:57:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.652	0.615	5.7	93	0.00
3 P	Chloromethane	0.791	0.693	12.4	92	0.00
4 C	Vinyl Chloride	0.671	0.639	4.8#	96	0.00
5 T	Bromomethane	0.363	0.326	10.2	94	0.00
6 T	Chloroethane	0.407	0.370	9.1	96	0.00
7 T	Trichlorofluoromethane	0.987	0.957	3.0	97	0.00
8 T	Diethyl Ether	0.380	0.362	4.7	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.586	0.548	6.5	96	0.00
10 T	Methyl Iodide	0.524	0.561	-7.1	110	0.00
11 T	Tert butyl alcohol	0.154	0.145	5.8	97	-0.01
12 CM	1,1-Dichloroethene	0.557	0.526	5.6#	95	0.00
13 T	Acrolein	0.183	0.158	13.7	93	0.00
14 T	Allyl chloride	1.193	1.109	7.0	95	0.00
15 T	Acrylonitrile	0.339	0.330	2.7	100	0.00
16 T	Acetone	0.375	0.309	17.6	89	0.00
17 T	Carbon Disulfide	1.344	1.266	5.8	93	0.00
18 T	Methyl Acetate	1.383	1.388	-0.4	102	0.00
19 T	Methyl tert-butyl Ether	2.216	2.092	5.6	98	0.00
20 T	Methylene Chloride	0.679	0.631	7.1	98	0.00
21 T	trans-1,2-Dichloroethene	0.610	0.579	5.1	95	0.00
22 T	Diisopropyl ether	2.336	2.187	6.4	98	0.00
23 T	Vinyl Acetate	1.653	1.594	3.6	98	0.00
24 P	1,1-Dichloroethane	1.249	1.192	4.6	98	0.00
25 T	2-Butanone	0.519	0.486	6.4	98	0.00
26 T	2,2-Dichloropropane	1.058	0.910	14.0	88	0.00
27 T	cis-1,2-Dichloroethene	0.733	0.696	5.0	98	0.00
28 T	Bromochloromethane	0.562	0.558	0.7	114	0.00
29 T	Tetrahydrofuran	0.326	0.313	4.0	101	0.00
30 C	Chloroform	1.280	1.216	5.0#	98	0.00
31 T	Cyclohexane	1.102	1.030	6.5	95	0.00
32 T	1,1,1-Trichloroethane	1.114	1.061	4.8	97	0.00
33 S	1,2-Dichloroethane-d4	0.872	0.868	0.5	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.337	0.333	1.2	101	0.00
36 T	1,1-Dichloropropene	0.532	0.493	7.3	95	0.00
37 T	Ethyl Acetate	0.555	0.537	3.2	99	0.00
38 T	Carbon Tetrachloride	0.504	0.477	5.4	96	0.00
39 T	Methylcyclohexane	0.627	0.571	8.9	93	0.00
40 TM	Benzene	1.489	1.380	7.3	97	0.00
41 T	Methacrylonitrile	0.330	0.308	6.7	97	0.00
42 TM	1,2-Dichloroethane	0.641	0.601	6.2	98	0.00
43 T	Isopropyl Acetate	0.968	0.923	4.6	99	0.00
44 TM	Trichloroethene	0.378	0.349	7.7	97	0.00
45 C	1,2-Dichloropropane	0.394	0.373	5.3#	97	0.00
46 T	Dibromomethane	0.270	0.256	5.2	97	0.00
47 T	Bromodichloromethane	0.522	0.504	3.4	98	0.00
48 T	Methyl methacrylate	0.478	0.466	2.5	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051123\
 Data File : VX035624.D
 Acq On : 11 May 2023 21:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 12 03:57:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.008	20.0	94	-0.01
50 S	Toluene-d8	1.246	1.219	2.2	99	0.00
51 T	4-Methyl-2-Pentanone	0.558	0.548	1.8	101	0.00
52 CM	Toluene	0.940	0.870	7.4#	97	0.00
53 T	t-1,3-Dichloropropene	0.567	0.550	3.0	95	0.00
54 T	cis-1,3-Dichloropropene	0.607	0.585	3.6	95	0.00
55 T	1,1,2-Trichloroethane	0.369	0.348	5.7	97	0.00
56 T	Ethyl methacrylate	0.598	0.581	2.8	100	0.00
57 T	1,3-Dichloropropane	0.669	0.626	6.4	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.304	0.308	-1.3	101	0.00
59 T	2-Hexanone	0.420	0.406	3.3	100	0.00
60 T	Dibromochloromethane	0.352	0.345	2.0	97	0.00
61 T	1,2-Dibromoethane	0.389	0.368	5.4	98	0.00
62 S	4-Bromofluorobenzene	0.502	0.495	1.4	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.337	0.319	5.3	102	0.00
65 PM	Chlorobenzene	1.091	1.022	6.3	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.365	5.4	96	0.00
67 C	Ethyl Benzene	2.053	1.911	6.9#	97	0.00
68 T	m/p-Xylenes	0.759	0.708	6.7	96	0.00
69 T	o-Xylene	0.753	0.698	7.3	96	0.00
70 T	Styrene	1.213	1.159	4.5	97	0.00
71 P	Bromoform	0.254	0.242	4.7	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	4.213	3.922	6.9	96	0.00
74 T	N-amyl acetate	1.964	1.854	5.6	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.367	1.265	7.5	99	0.00
76 T	1,2,3-Trichloropropane	1.223	1.088	11.0	99	0.00
77 T	Bromobenzene	0.956	0.876	8.4	98	0.00
78 T	n-propylbenzene	4.953	4.616	6.8	96	0.00
79 T	2-Chlorotoluene	3.101	2.848	8.2	97	0.00
80 T	1,3,5-Trimethylbenzene	3.581	3.315	7.4	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.379	0.354	6.6	94	0.00
82 T	4-Chlorotoluene	3.571	3.284	8.0	97	0.00
83 T	tert-Butylbenzene	3.481	3.205	7.9	97	0.00
84 T	1,2,4-Trimethylbenzene	3.596	3.346	7.0	97	0.00
85 T	sec-Butylbenzene	4.370	4.042	7.5	95	0.00
86 T	p-Isopropyltoluene	3.593	3.375	6.1	95	0.00
87 T	1,3-Dichlorobenzene	1.817	1.641	9.7	95	0.00
88 T	1,4-Dichlorobenzene	1.834	1.655	9.8	94	0.00
89 T	n-Butylbenzene	3.268	3.022	7.5	93	0.00
90 T	Hexachloroethane	0.545	0.511	6.2	93	0.00
91 T	1,2-Dichlorobenzene	1.779	1.611	9.4	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.304	0.290	4.6	100	0.00
93 T	1,2,4-Trichlorobenzene	1.071	0.993	7.3	95	0.00
94 T	Hexachlorobutadiene	0.425	0.385	9.4	93	0.00
95 T	Naphthalene	3.668	3.484	5.0	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051123\
Data File : VX035624.D
Acq On : 11 May 2023 21:43
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 12 03:57:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.053	0.980	6.9	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6