



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051123\
Data File : VX035599.D
Acq On : 11 May 2023 10:10
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Title : SW846 8260
Last Update : Thu May 11 13:57:54 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.6	14462	PASS
75	95	30	60	54.5	38203	PASS
95	95	100	100	100.0	70040	PASS
96	95	5	9	6.9	4828	PASS
173	174	0.00	2	0.7	303	PASS
174	95	50	100	65.8	46104	PASS
175	174	5	9	7.7	3531	PASS
176	174	95	101	95.8	44157	PASS
177	176	5	9	6.6	2912	PASS