

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051124\
 Data File : VX041455.D
 Acq On : 10 May 2024 13:41
 Operator : JC/MD
 Sample : VX0511WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0511WBL01

Quant Time: May 11 01:33:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	156845	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	229759	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217088	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86709	50.782	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.560%	
35) Dibromofluoromethane	5.385	113	71921	46.192	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 92.380%	
50) Toluene-d8	8.647	98	267245	54.229	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.460%	
62) 4-Bromofluorobenzene	11.079	95	98944	51.866	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.740%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051124\
Data File : VX041455.D
Acq On : 10 May 2024 13:41
Operator : JC/MD
Sample : VX0511WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0511WBL01

Quant Time: May 11 01:33:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

