

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051124\
 Data File : VX041459.D
 Acq On : 10 May 2024 15:25
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: May 11 01:35:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	155666	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	229774	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213206	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	110962	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86488	51.036	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.080%
35) Dibromofluoromethane	5.385	113	70910	45.539	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.080%
50) Toluene-d8	8.646	98	268006	54.380	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.760%
62) 4-Bromofluorobenzene	11.079	95	93105	48.802	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051124\
Data File : VX041459.D
Acq On : 10 May 2024 15:25
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: May 11 01:35:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

