

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051124\
 Data File : VX041464.D
 Acq On : 10 May 2024 17:20
 Operator : JC/MD
 Sample : P2448-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: May 11 01:37:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	177123	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	272462	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	257492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125068	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	105965	54.955	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 109.900%		
35) Dibromofluoromethane	5.385	113	86063	46.611	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 93.220%		
50) Toluene-d8	8.647	98	313136	53.582	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 107.160%		
62) 4-Bromofluorobenzene	11.079	95	117138	51.779	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 103.560%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051124\
Data File : VX041464.D
Acq On : 10 May 2024 17:20
Operator : JC/MD
Sample : P2448-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: May 11 01:37:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

