

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051124\
 Data File : VX041468.D
 Acq On : 10 May 2024 18:53
 Operator : JC/MD
 Sample : P2455-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-02B-66-051024

Quant Time: May 11 01:38:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	159281	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	236018	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	118411	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91161	52.573	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.140%			
35) Dibromofluoromethane	5.385	113	73781	46.130	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 92.260%			
50) Toluene-d8	8.646	98	269001	53.138	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 106.280%			
62) 4-Bromofluorobenzene	11.079	95	91353	46.617	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 93.240%			
Target Compounds						
4) Vinyl Chloride	1.373	62	4916	3.808	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.318	101	18185	14.533	ug/l	90
12) 1,1-Dichloroethene	2.312	96	207097	167.785	ug/l	87
16) Acetone	2.391	43	10030	9.371	ug/l	93
21) trans-1,2-Dichloroethene	3.086	96	7613	5.519	ug/l	91
24) 1,1-Dichloroethane	3.611	63	43723	17.610	ug/l	97
27) cis-1,2-Dichloroethene	4.483	96	469833	287.673	ug/l	82
30) Chloroform	5.092	83	3458	1.339	ug/l	93
40) Benzene	6.043	78	4264	0.723	ug/l	94
42) 1,2-Dichloroethane	6.086	62	22846	10.573	ug/l	94
44) Trichloroethene	7.128	130	4899824	3147.268	ug/l	88
64) Tetrachloroethene	9.274	164	67261	41.061	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051124\
Data File : VX041468.D
Acq On : 10 May 2024 18:53
Operator : JC/MD
Sample : P2455-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-02B-66-051024

Quant Time: May 11 01:38:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

