

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051124\
 Data File : VX041469.D
 Acq On : 10 May 2024 19:16
 Operator : JC/MD
 Sample : P2445-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-04B-92.5-050924

Quant Time: May 11 01:39:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	157340	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	233308	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223685	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109904	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89807	52.431	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	104.860%	
35) Dibromofluoromethane	5.391	113	71723	45.364	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	90.720%	
50) Toluene-d8	8.647	98	267532	53.462	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	106.920%	
62) 4-Bromofluorobenzene	11.079	95	86317	44.559	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	89.120%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	1174	0.950	ug/l	87
12) 1,1-Dichloroethene	2.313	96	521	0.427	ug/l	86
16) Acetone	2.386	43	4452	2.022	ug/l	98
30) Chloroform	5.105	83	1984	0.778	ug/l	90
44) Trichloroethene	7.123	130	19386	12.597	ug/l	90
64) Tetrachloroethene	9.281	164	383	0.236	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051124\
Data File : VX041469.D
Acq On : 10 May 2024 19:16
Operator : JC/MD
Sample : P2445-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-04B-92.5-050924

Quant Time: May 11 01:39:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

