

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
 Data File : VX028642.D
 Acq On : 12 May 2022 16:44
 Operator : JC/MD
 Sample : VX0512WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0512WBL01

Quant Time: May 13 02:47:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	247149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	456499	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	428206	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	192308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	176281	52.033	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.060%	
35) Dibromofluoromethane	5.385	113	150875	50.444	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.880%	
50) Toluene-d8	8.646	98	548418	48.635	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.260%	
62) 4-Bromofluorobenzene	11.079	95	200703	47.653	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
Data File : VX028642.D
Acq On : 12 May 2022 16:44
Operator : JC/MD
Sample : VX0512WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0512WBL01

Quant Time: May 13 02:47:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

