

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
 Data File : VX028647.D
 Acq On : 12 May 2022 18:49
 Operator : JC/MD
 Sample : N2799-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: May 13 02:48:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	218311	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	410200	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	391001	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	178524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	163100	54.502	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.000%	
35) Dibromofluoromethane	5.391	113	137932	51.322	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.640%	
50) Toluene-d8	8.653	98	500905	49.435	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.860%	
62) 4-Bromofluorobenzene	11.085	95	183108	48.382	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.760%	
Target Compounds						
						Qvalue
16) Acetone	2.373	43	7744	6.030	ug/l	93
17) Carbon Disulfide	2.520	76	7431	1.269	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
Data File : VX028647.D
Acq On : 12 May 2022 18:49
Operator : JC/MD
Sample : N2799-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: May 13 02:48:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

