

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
 Data File : VX028658.D
 Acq On : 12 May 2022 23:05
 Operator : JC/MD
 Sample : N2811-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: May 13 02:50:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	220903	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	411056	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	392925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176530	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	162161	53.553	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.100%	
35) Dibromofluoromethane	5.391	113	135886	50.456	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.920%	
50) Toluene-d8	8.653	98	500735	49.315	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.640%	
62) 4-Bromofluorobenzene	11.079	95	184578	48.669	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.340%	
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	3270	2.516	ug/l	91
20) Methylene Chloride	2.788	84	701	0.257	ug/l #	87
64) Tetrachloroethene	9.281	164	1145	0.386	ug/l	94
68) m/p-Xylenes	10.305	106	1267	0.218	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
Data File : VX028658.D
Acq On : 12 May 2022 23:05
Operator : JC/MD
Sample : N2811-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: May 13 02:50:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

