

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
 Data File : VX028659.D
 Acq On : 12 May 2022 23:29
 Operator : JC/MD
 Sample : N2811-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: May 13 02:50:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	220182	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	409320	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	391815	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176691	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	162617	53.879	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 107.760%		
35) Dibromofluoromethane	5.391	113	134462	50.139	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 100.280%		
50) Toluene-d8	8.653	98	500056	49.457	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 98.920%		
62) 4-Bromofluorobenzene	11.079	95	184268	48.793	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 97.580%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2389	1.844	ug/l	90
27) cis-1,2-Dichloroethene	4.489	96	1983	0.661	ug/l	97
30) Chloroform	5.105	83	2458	0.500	ug/l	94
64) Tetrachloroethene	9.275	164	15142	5.118	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
Data File : VX028659.D
Acq On : 12 May 2022 23:29
Operator : JC/MD
Sample : N2811-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: May 13 02:50:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

