

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
 Data File : VX028663.D
 Acq On : 13 May 2022 01:02
 Operator : JC/MD
 Sample : N2775-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-35S

Quant Time: May 13 02:51:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	237506	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	439579	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	434986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	192925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	174484	53.594	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 107.180%		
35) Dibromofluoromethane	5.391	113	145827	50.633	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 101.260%		
50) Toluene-d8	8.653	98	545304	50.220	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 100.440%		
62) 4-Bromofluorobenzene	11.079	95	200131	49.346	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 98.700%		
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	249590	178.629	ug/l	91
20) Methylene Chloride	2.788	84	25949	8.852	ug/l	90
25) 2-Butanone	4.562	43	34606	15.779	ug/l	96
43) Isopropyl Acetate	6.348	43	23133	3.273	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
Data File : VX028663.D
Acq On : 13 May 2022 01:02
Operator : JC/MD
Sample : N2775-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-35S

Quant Time: May 13 02:51:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

