

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
 Data File : VX028664.D
 Acq On : 13 May 2022 01:25
 Operator : JC/MD
 Sample : N2775-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-35D

Quant Time: May 13 02:51:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	237638	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	448848	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	435805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	175484	53.871	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	107.740%	
35) Dibromofluoromethane	5.391	113	146878	49.945	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	99.900%	
50) Toluene-d8	8.653	98	551081	49.704	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	99.400%	
62) 4-Bromofluorobenzene	11.079	95	203529	49.147	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	98.300%	
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	168155	120.280	ug/l	91
20) Methylene Chloride	2.794	84	19145	6.527	ug/l	93
25) 2-Butanone	4.562	43	34047	15.515	ug/l	93
43) Isopropyl Acetate	6.342	43	22905	3.174	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051222\
Data File : VX028664.D
Acq On : 13 May 2022 01:25
Operator : JC/MD
Sample : N2775-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-35D

Quant Time: May 13 02:51:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

