

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051223\
 Data File : VX035633.D
 Acq On : 12 May 2023 13:21
 Operator : JC/MD
 Sample : 02755-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A5560

Quant Time: May 15 01:55:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	184965	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	334353	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	306327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132347	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	154528	47.889	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.780%		
35) Dibromofluoromethane	5.385	113	110242	48.891	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.780%		
50) Toluene-d8	8.647	98	405014	48.621	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.240%		
62) 4-Bromofluorobenzene	11.079	95	164336	48.997	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.000%		
Target Compounds						
						Qvalue
16) Acetone	2.398	43	3778	2.724	ug/l	89
17) Carbon Disulfide	2.514	76	1234	0.248	ug/l	100
20) Methylene Chloride	2.788	84	1018	0.405	ug/l	# 83
44) Trichloroethene	7.135	130	649	0.257	ug/l	68
51) 4-Methyl-2-Pentanone	8.573	43	978	0.262	ug/l	# 83
52) Toluene	8.720	92	71586	11.390	ug/l	98
64) Tetrachloroethene	9.275	164	613	0.297	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051223\
Data File : VX035633.D
Acq On : 12 May 2023 13:21
Operator : JC/MD
Sample : 02755-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A5560

Quant Time: May 15 01:55:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

