

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051223\
 Data File : VX035636.D
 Acq On : 12 May 2023 14:31
 Operator : JC/MD
 Sample : 02777-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: May 15 01:56:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	182707	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	331056	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	298513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137707	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	152033	47.698	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.400%
35) Dibromofluoromethane	5.385	113	106202	47.568	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.140%
50) Toluene-d8	8.647	98	401611	48.693	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.380%
62) 4-Bromofluorobenzene	11.079	95	164546	49.548	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.100%
Target Compounds						
16) Acetone	2.392	43	1742	1.272	ug/l	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051223\
Data File : VX035636.D
Acq On : 12 May 2023 14:31
Operator : JC/MD
Sample : 02777-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: May 15 01:56:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

