

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
 Data File : VX046131.D
 Acq On : 12 May 2025 12:26
 Operator : JC/MD
 Sample : Q1983-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-636-COMP-01

Quant Time: May 13 03:16:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	69835	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	139270	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	132799	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70525	54.169	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.340%		
35) Dibromofluoromethane	5.379	113	52384	52.233	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.460%		
50) Toluene-d8	8.647	98	178099	51.308	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.620%		
62) 4-Bromofluorobenzene	11.079	95	67060	50.365	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 100.740%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	18441	35.326	ug/l	97
18) Methyl Acetate	2.709	43	1311	1.082	ug/l	96
20) Methylene Chloride	2.782	84	2691	2.690	ug/l #	90
43) Isopropyl Acetate	6.348	43	9025	3.553	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
Data File : VX046131.D
Acq On : 12 May 2025 12:26
Operator : JC/MD
Sample : Q1983-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-636-COMP-01

Quant Time: May 13 03:16:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

