

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
 Data File : VX046132.D
 Acq On : 12 May 2025 12:49
 Operator : JC/MD
 Sample : Q1983-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-636-COMP-02

Quant Time: May 13 03:16:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	74922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	150365	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138880	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75584	54.113	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.220%			
35) Dibromofluoromethane	5.385	113	55466	51.225	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.460%			
50) Toluene-d8	8.646	98	188300	50.245	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.480%			
62) 4-Bromofluorobenzene	11.079	95	71028	49.409	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 98.820%			
Target Compounds					Qvalue	
16) Acetone	2.379	43	18167	32.438	ug/l	98
18) Methyl Acetate	2.702	43	1711	1.317	ug/l	95
20) Methylene Chloride	2.788	84	2635	2.455	ug/l	95
43) Isopropyl Acetate	6.342	43	10979	4.004	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
Data File : VX046132.D
Acq On : 12 May 2025 12:49
Operator : JC/MD
Sample : Q1983-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-636-COMP-02

Quant Time: May 13 03:16:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

