

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
 Data File : VX046133.D
 Acq On : 12 May 2025 13:12
 Operator : JC/MD
 Sample : Q1983-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-636-COMP-03

Quant Time: May 13 03:17:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66694	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135660	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	58030	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67171	54.022	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.040%		
35) Dibromofluoromethane	5.379	113	49935	51.116	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.240%		
50) Toluene-d8	8.647	98	173440	51.296	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.600%		
62) 4-Bromofluorobenzene	11.079	95	67308	51.896	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 103.800%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	19585	39.285	ug/l	96
18) Methyl Acetate	2.703	43	1811	1.565	ug/l	92
20) Methylene Chloride	2.782	84	2513	2.630	ug/l	96
43) Isopropyl Acetate	6.348	43	10574	4.274	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
Data File : VX046133.D
Acq On : 12 May 2025 13:12
Operator : JC/MD
Sample : Q1983-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-636-COMP-03

Quant Time: May 13 03:17:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

