

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
 Data File : VX046139.D
 Acq On : 12 May 2025 15:32
 Operator : JC/MD
 Sample : Q1983-54
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-636-COMP-09

Quant Time: May 13 03:19:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	134215	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68759	55.316	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.640%
35) Dibromofluoromethane	5.385	113	49472	51.187	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.380%
50) Toluene-d8	8.647	98	172797	51.656	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.320%
62) 4-Bromofluorobenzene	11.079	95	65121	50.751	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.500%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	18162	36.441	ug/l	99
18) Methyl Acetate	2.709	43	2269	1.962	ug/l	93
20) Methylene Chloride	2.782	84	2450	2.565	ug/l	86
43) Isopropyl Acetate	6.342	43	10447	4.268	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
Data File : VX046139.D
Acq On : 12 May 2025 15:32
Operator : JC/MD
Sample : Q1983-54
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-636-COMP-09

Quant Time: May 13 03:19:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

