

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
 Data File : VX046141.D
 Acq On : 12 May 2025 16:19
 Operator : JC/MD
 Sample : Q1986-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-9

Quant Time: May 13 03:20:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65341	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131886	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	127831	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68734	56.424	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125		Recovery	= 112.840%	
35) Dibromofluoromethane	5.379	113	49513	52.134	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124		Recovery	= 104.260%	
50) Toluene-d8	8.647	98	171223	52.089	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113		Recovery	= 104.180%	
62) 4-Bromofluorobenzene	11.079	95	63745	50.556	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121		Recovery	= 101.120%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	18609	38.100	ug/l	92
18) Methyl Acetate	2.709	43	2186	1.929	ug/l	97
20) Methylene Chloride	2.788	84	2587	2.764	ug/l #	96
43) Isopropyl Acetate	6.355	43	10868	4.519	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
Data File : VX046141.D
Acq On : 12 May 2025 16:19
Operator : JC/MD
Sample : Q1986-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-9

Quant Time: May 13 03:20:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

