

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
 Data File : VX046142.D
 Acq On : 12 May 2025 16:42
 Operator : JC/MD
 Sample : Q1986-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-10

Quant Time: May 13 03:20:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65301	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132415	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55187	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67700	55.609	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.220%
35) Dibromofluoromethane	5.385	113	48925	51.309	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.620%
50) Toluene-d8	8.647	98	169707	51.422	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.840%
62) 4-Bromofluorobenzene	11.079	95	64529	50.973	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.940%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	19471	39.889	ug/l	94
18) Methyl Acetate	2.703	43	2074	1.831	ug/l	98
20) Methylene Chloride	2.782	84	2610	2.790	ug/l	84
43) Isopropyl Acetate	6.348	43	10438	4.323	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
Data File : VX046142.D
Acq On : 12 May 2025 16:42
Operator : JC/MD
Sample : Q1986-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-10

Quant Time: May 13 03:20:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

