

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
 Data File : VX046144.D
 Acq On : 12 May 2025 17:29
 Operator : JC/MD
 Sample : Q1986-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-218

Quant Time: May 13 03:21:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62310	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124531	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121494	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51299	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63033	54.261	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.520%
35) Dibromofluoromethane	5.385	113	46779	52.165	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.320%
50) Toluene-d8	8.647	98	160923	51.847	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.700%
62) 4-Bromofluorobenzene	11.079	95	61480	51.639	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.280%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	18925	40.632	ug/l	97
18) Methyl Acetate	2.703	43	2600	2.406	ug/l #	85
20) Methylene Chloride	2.788	84	2504	2.805	ug/l	93
43) Isopropyl Acetate	6.348	43	10787	4.750	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
Data File : VX046144.D
Acq On : 12 May 2025 17:29
Operator : JC/MD
Sample : Q1986-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-218

Quant Time: May 13 03:21:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

