

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
 Data File : VX046145.D
 Acq On : 12 May 2025 17:52
 Operator : JC/MD
 Sample : Q2004-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-220

Quant Time: May 13 03:21:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131713	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51911	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67227	55.827	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.660%
35) Dibromofluoromethane	5.379	113	49035	51.699	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.400%
50) Toluene-d8	8.647	98	167185	50.928	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.860%
62) 4-Bromofluorobenzene	11.079	95	63162	50.159	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.320%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	19752	40.909	ug/l	99
18) Methyl Acetate	2.709	43	2442	2.180	ug/l #	86
20) Methylene Chloride	2.782	84	3585	3.875	ug/l #	96
43) Isopropyl Acetate	6.355	43	9926	4.132	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
Data File : VX046145.D
Acq On : 12 May 2025 17:52
Operator : JC/MD
Sample : Q2004-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-220

Quant Time: May 13 03:21:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

