

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
 Data File : VX046146.D
 Acq On : 12 May 2025 18:16
 Operator : JC/MD
 Sample : Q2004-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2811

Quant Time: May 13 03:22:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122025	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	117034	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47598	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62733	54.503	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.000%		
35) Dibromofluoromethane	5.379	113	45483	51.761	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.520%		
50) Toluene-d8	8.647	98	158363	52.070	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 104.140%		
62) 4-Bromofluorobenzene	11.079	95	57837	49.577	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 99.160%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	20311	44.011	ug/l	98
18) Methyl Acetate	2.703	43	2010	1.877	ug/l #	90
20) Methylene Chloride	2.782	84	3689	4.171	ug/l	87
43) Isopropyl Acetate	6.355	43	9932	4.463	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
Data File : VX046146.D
Acq On : 12 May 2025 18:16
Operator : JC/MD
Sample : Q2004-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2811

Quant Time: May 13 03:22:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

