

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
 Data File : VX046149.D
 Acq On : 12 May 2025 19:26
 Operator : JC/MD
 Sample : Q2004-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SOIL-STOCK

Quant Time: May 13 03:23:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	60204	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	121882	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	118844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51751	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	61646	54.924	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.840%	
35) Dibromofluoromethane	5.385	113	44850	51.101	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.200%	
50) Toluene-d8	8.647	98	157367	51.804	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.600%	
62) 4-Bromofluorobenzene	11.079	95	60392	51.828	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.660%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	26524	58.939	ug/l	98
18) Methyl Acetate	2.703	43	2319	2.221	ug/l	99
20) Methylene Chloride	2.794	84	2408	2.792	ug/l #	58
43) Isopropyl Acetate	6.342	43	13378	6.019	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
Data File : VX046149.D
Acq On : 12 May 2025 19:26
Operator : JC/MD
Sample : Q2004-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SOIL-STOCK

Quant Time: May 13 03:23:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

