

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
 Data File : VX046150.D
 Acq On : 12 May 2025 19:49
 Operator : JC/MD
 Sample : Q2004-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STONE-STOCK

Quant Time: May 13 03:23:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	61700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	123008	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	117848	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63781	55.448	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.900%	
35) Dibromofluoromethane	5.379	113	46172	52.125	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	104.260%	
50) Toluene-d8	8.647	98	157098	51.242	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.480%	
62) 4-Bromofluorobenzene	11.079	95	57927	49.257	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	98.520%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	20945	45.413	ug/l	98
18) Methyl Acetate	2.709	43	2052	1.917	ug/l #	86
20) Methylene Chloride	2.782	84	3633	4.111	ug/l	96
43) Isopropyl Acetate	6.355	43	4061	1.810	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051225\
Data File : VX046150.D
Acq On : 12 May 2025 19:49
Operator : JC/MD
Sample : Q2004-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STONE-STOCK

Quant Time: May 13 03:23:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

