

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051319\
 Data File : VX009538.D
 Acq On : 13 May 2019 13:12
 Operator : JC/SP
 Sample : K2799-03DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3DL

Quant Time: May 14 04:04:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	195482	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	306775	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	261814	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	103481	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	124516	46.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.32%	
35) Dibromofluoromethane	5.50	113	93070	48.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.52%	
50) Toluene-d8	8.71	98	383266	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
62) 4-Bromofluorobenzene	11.13	95	120404	42.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.12%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	33851	14.115	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	119649	50.393	ug/l	88
44) Trichloroethene	7.21	130	7333	3.438	ug/l	88
64) Tetrachloroethene	9.34	164	1160	0.646	ug/l	87
91) 1,2-Dichlorobenzene	12.39	146	1583	0.467	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051319\
Data File : VX009538.D
Acq On : 13 May 2019 13:12
Operator : JC/SP
Sample : K2799-03DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3DL

Quant Time: May 14 04:04:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

