

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051319\
 Data File : VX009549.D
 Acq On : 13 May 2019 17:28
 Operator : JC/SP
 Sample : K2840-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KR-04 9-19

Quant Time: May 14 04:49:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	188490	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	300646	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	262473	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	114190	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	123076	47.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.66%	
35) Dibromofluoromethane	5.49	113	91199	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
50) Toluene-d8	8.71	98	374687	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
62) 4-Bromofluorobenzene	11.13	95	124647	44.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.92%	
Target Compounds						
64) Tetrachloroethene	9.34	164	1910	1.061	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051319\
Data File : VX009549.D
Acq On : 13 May 2019 17:28
Operator : JC/SP
Sample : K2840-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KR-04_9-19

Quant Time: May 14 04:49:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

